



REVISIONS					
REV	ECO No.	DESCRIPTION	BY	APPROVED	DATE

4 GRAVITY DRAIN HAS A 1-1/2 NOMINAL PIPE PROTRUDING FROM FINISHED FLOOR
NOTE: PLACE TRAP IN LINE TO CONFORM WITH LOCAL CODES, SUPPLIED BY CONTRACTOR

3 PLUMBING CENTRAL VACUUM TO THE UTILITY CENTER SHOULD BE SPECIFIED BY THE CENTRAL VACUUM SUPPLIER AND TERMINATED IN UTILITY CENTER WITH 3/4 O.D. TUBE PERPENDICULAR TO THE FLOOR, SIMILAR TO THE DRAIN CONNECTION

2 MAKE SURE ALL PLUMBING CONFORMS TO PREVAILING LOCAL CODES

1 LOCAL REGULATIONS PROVIDE THAT LICENSED PLUMBERS & ELECTRICIANS SHALL INSTALL UTILITIES

NOTES:

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ARE:

DECIMALS		FRACTIONS	
.X = ±.030		X/X = ±1/32	
.XX = ±.010			
.XXX = ±.005		ANGULAR	±1°

TPC ADVANCED TECHNOLOGY
19388 E Rowland PL Suite 296
City of Industry, CA 91748
PH 626-810-4327
FAX 626-810-4402
www.tpcadvanced.com

TITLE: TEMPLATE, UTILITY CENTER

APPV	DATE:
DRAWN:	
CHECKED:	
ENGGR:	

MATERIAL: | PART No.: | REV: —

FINISH: | SIZE: D | SCALE: FULL | SHEET 1 OF 1