## CITY OF LOS ANGELES

**CALIFORNIA** 





**BOARD OF** 

**BUILDING AND SAFETY** 

COMMISSIONERS

JOSELYN GEAGA-ROSENTHAL
VICE PRESIDENT

JACOB STEVENS MOISES ROSALES NANCY YAP



KAREN BASS MAYOR OSAMA YOUNAN, P.E. GENERAL MANAGER SUPERINTENDENT OF BUILDING

> JOHN WEIGHT EXECUTIVE OFFICER

**Scott Beckley** 

TPC Advanced Technology, Inc.

851 S. Lawson St.

City of Industry, CA 91748

Telephone: 626-810-4337

File Number: M-030007 Effective Date: 06/01/2023 Expiration Date: 06/01/2024

Manufacturer: TPC Advanced Technology,

Inc.

Item: Dental Units

Models:

Cosmetica, Mirage & Bella - TPC Models - 2000, 2015, 2350, 2375, 2530, 2630, 2635, 2400, 2450, 2500, 2550 and 2600 Redland Models: 3500, 3550, MC500, MC501, MS-3000, and MR-4000.

## PLEASE REFER TO THE APPLICATION NUMBER ON ALL CORRESPONDENCE, PHONE CALLS & SAMPLES

The item processed under the above file number has been tested, examined and found to comply with applicable Los Angeles Plumbing Code (LAPC) and/or Los Angeles Mechanical Code (LAMC) and is hereby approved unless revoked for cause.

This letter or the letterhead copy may not be reproduced for use in labeling or advertising, but copies of this letter may be used within your organization and shown as evidence of approval.

## CONDITIONS OF APPROVAL:

- 1. An approved pressure vacuum breaker shall be installed on each potable water inlet at least 12-inches above all downstream piping and flood-level rim receptor per Section 94.603.0 of the LAPC, 2023 Edition.
- 2. Each drain line shall be routed without dips or sags to terminate 1-inch or two-pipe sizes, whichever is greater, above the flood-level rim of an approved indirect waste receptor per Section 94.801.0 of the LAPC, 2023 Edition or shall be connected to an approved vacuum pump system.

  Note: The hand piece has a built-in check valve and the cuspidor bowl rinse nozzle is not less than 1-inch above the bowl.

Each item shall have the name or trademark of manufacturer and model number where it will be visible for inspection.

In order to renew this approval, an application and reexamination fee must be submitted on or before the expiration date shown above. Arrangement shall be made to remove any test samples remaining in the Laboratory. Samples remaining after 21 days from the date of this notice will be shipped collect by the most convenient carrier or disposed of in accordance with applicable regulation.

Fuad Malki

LADBS G-5 (Rev. 05/30/2023)

Mechanical Testing Laboratory, Test Engineer